

OFFICE OF THE SECRETARY OF STATE
DENNIS RICHARDSON
SECRETARY OF STATE

LESLIE CUMMINGS
DEPUTY SECRETARY OF STATE



ARCHIVES DIVISION
MARY BETH HERKERT
DIRECTOR

800 SUMMER STREET NE
SALEM, OR 97310
503-373-0701

STATUTORY MINOR CORRECTION

PH 19-2018

CHAPTER 333
OREGON HEALTH AUTHORITY
PUBLIC HEALTH DIVISION

FILED

02/01/2018 12:11 PM
ARCHIVES DIVISION
SECRETARY OF STATE
& LEGISLATIVE COUNSEL

FILING CAPTION: Radiation Protection Services - Statutory minor correction to rule titles

CONTACT: Brittany Hall
503-449-9808
publichealth.rules@state.or.us

800 NE Oregon St. Suite 930
Portland, OR 97232

Filed By:
Brittany Hall
Rules Coordinator

AGENCY ATTESTS THE FOLLOWING CHANGES HAVE BEEN MADE, ACCORDING TO ORS 183.335(7):

Changing a rule title according to ORS 183.360(2)(a)

AMEND: 333-116-0310

RULE TITLE: Possession of Survey Instrument

RULE SUMMARY: Rule titles in OAR 333-116-0150 through 333-116-1030 are being amended to clear up confusion as to what they apply to. These rule titles were assigned specific sub-categories when subdivisions of rules were phased out during the implementation of the Oregon Administrative Rules Database (OARD). However, rules impact multiple medical modalities causing confusion to licensees and stakeholders. The statutory minor correction removes the sub-category from rule title.

RULE TEXT:

A licensee authorized to use radioactive material for uptake, dilution and excretion studies must have in its possession a portable radiation detection survey instrument capable of detecting dose rates over the range one Sv (0.1 mrem) per hour to one mSv (100 mrem) per hour. The instrument must be operable and calibrated in accordance with OAR 333-116-0170.

STATUTORY/OTHER AUTHORITY: ORS 453.635

STATUTES/OTHER IMPLEMENTED: ORS 453.605 - 453.807